

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Electronic measuring equipment
Name and address of the applicant	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19 74385 Pleidelsheim GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	Rated Voltage: 100-120 V / 200-240 V Rated frequency: 50/60 Hz Rated power: 600 VA Protection class: I Air pressure: 6-8 bar Stripe projector: Risk class 2 (acc. IEC 62471)
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	mµFocus S
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713190849

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 095184 0031 Rev. 00

Date, 2021-02-17



(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service