



## CERTIFICATE

No. U8 095184 0028 Rev. 00

Holder of Certificate:

## E. ZOLLER GmbH & Co. KG

Gottlieb-Daimler-Str. 19 74385 Pleidelsheim GERMANY

Production Facility(ies): 095184

**Certification Mark:** 



**Product:** 

Measuring device

Model(s):

venturion, smartCheck, phoenix, millCheck, hobCheck, redomatic, reamCheck, scanTast

100-120 V / 200-240 V Rated Voltage: Parameters: Rated frequency: 50/60 Hz 600 VA Rated power: Protection class: 6-8 bar Air pressure:

## Tested according to:

UL 61010-1:2012/R:2018-11 CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.:

028-713173838-000

Date, 2020-08-11

Kist

(Ralph Fischer)